



TEST REPORT



Page 1 of 4 Pages

No.:WT204044871

Sample Description: Medical surgical mask

Model/Specification/Grade: Flat form(ear loop) 17.5cm×9.5cm

Applicant: Hunan Nun Medical Supplies Co., Ltd.

Applicant Address: Refer To Next Pages

Date of Receipt: 2020-05-28

Test Period: 2020-05-28 to 2020-06-22

Test Location: Longhua Experimental Base

Shenzhen Academy of
Metrology & Quality Inspection
(Stamp)

Approved by:

邓海英

Issue Date: 2020-06-22

Signature:



TEST REPORT

No.: WT204044871

Page 2 of 4 Pages

Sample Information:

Sample Description: Medical surgical mask
Trade Mark: 努恩 NUN
Model/Specification/Grade: Flat form(ear loop) 17.5cm×9.5cm
Serial/Batch No. of Sample: 20200502
Manufactured Date: Date: 2020-05-10
Manufacturer: Hunan Nun Medical Supplies Co., Ltd.
Manufacturer Address: Refer To Next Pages
Sample Quantity: 200PCS
Sample Description before Testing: Normal.

Client Information:

Applicant: Hunan Nun Medical Supplies Co., Ltd.
Applicant Address: Refer To Next Pages
Applicant Telephone: 13470462222
Applicant Post Code: -----

Test Information:

Date of Receipt: 2020-05-28
Applicant No.: 8298492
Environment Condition: (18~25) °C (30~70) %RH
Sampling Method: Delivered by Applicant
Judgment Basis: EN 14683:2019+AC:2019
Test Standard: ISO 22609:2004 and other method standards refer to next pages.

Test Conclusion:

Test result refer to next pages.

Tested by: 林月明

Checked by: 王燕归



TEST REPORT

No.: WT204044871

Page 3 of 4 Pages

Test Item	Requirement	Test Result	Conclusion
Bacterial filtration efficiency (BFE) (%)	Type IIR ≥ 98	Sample 1: 99.9 Sample 2: 100.0 Sample 3: 100.0 Sample 4: 100.0 Sample 5: 100.0	PASS
EN 14683:2019+AC:2019, Annex B			
Differential pressure (Pa/cm ²)	Type IIR < 60	Average value: <u>45.2</u> Flow rate: 8L/min	PASS
EN 14683:2019+AC:2019, Annex C			
Splash resistance pressure (kPa)	Type IIR ≥ 16.0	AQL4.0%, the highest corresponding blood pressure: <u>21.3</u>	PASS
ISO 22609:2004			
Microbial cleanliness (cfu/g)	Type IIR ≤ 30	Sample 1: 17.2 Sample 2: 15.8 Sample 3: 6.8 Sample 4: 9.1 Sample 5: 14.9	PASS
EN 14683:2019+AC:2019			

Note:

1. Applicant address、Manufacturer address:

#404 and #504, No. 3 Plant, Xingsha International Enterprise Center, No.68 Luositang Road, National Economic & Technical Development Zone, Changsha, Hunan, China.



TEST REPORT

No. : WT204044871

Page 4 of 4 Pages



END OF REPORT